# Audit procedure and plan; Round robins








## Objectives of the Audit

According to the Measurement Guidelines for NO<sub>x</sub> and VOCs<sup>\*</sup>, ACTRIS (2018), section 7.2 "Audit Procedures":



2. Check of the **conformity** of measurement of test gases with targeted values within the **DQOs**.







### Preconditions for a CiGas Audit

### NF has successfully completed Labelling step 1a

- Conformity check of the reported instrument setup based on the questionnaire
- ➤ DG has confirmed labelling step 1a (TC recommendation, RICommapproval, provision of a committment letter)
- Upgrades have been fully implemented
- NF submits data to the ACTRIS data base












### Audit workflow

On-site audit: 3 Days





## Audit procedures

#### 1. Instrument and installations

- All parts of the sampling and instrument set-up
- Calibration and zero gas systems
- Overall equipment of the station

#### 2. Training and operation

- Training of operators and instructions at the station (SOP)
- Measurement of intercomparison through RR cylinders (VOC)

#### 3. Documentation

- QA and QC data
- Logbooks
- SOP

#### 4. Evaluation of data

- Calibration, zero gas, target gas, and standard addition data
- Data delivery
- Results from intercomparison exercises (RR cylinders)
- Uncertainty evaluation










### Round-Robin Procedure for NMHCs



1 x NPL NMHC Standard (4 ppb in N<sub>2</sub>) 2 x NMHC Mixtures in amb. Air







ACTRIS CiGas Community work





## Audit for NO<sub>x</sub>

- Zero evaluation of station instruments
- Dry calibration (40 min cycle each for zero, NO, NO<sub>2</sub> via GPT)
- Humid calibration (40 min cycle each for zero, NO, NO<sub>2</sub> via GPT)
- Humidity dependence NO+O<sub>3</sub> of NO<sub>x</sub> instruments
- Influence of ozone through the inlet sampling line of the station (via NO+O<sub>3</sub>)
- Ambient air measurement (overnight)










## Audit for NO<sub>x</sub>

MGDS  $\rightarrow$  FZJ system to introduce dry/humidified air, calibration gas and O<sub>3</sub> NO<sub>x</sub>  $\rightarrow$  Airyx ICAD-NO<sub>x</sub> monitor; Airyx ICAD-HONO/NO<sub>2</sub> monitor Picarro G2301  $\rightarrow$  H<sub>2</sub>O, CO<sub>2</sub> 2BTech 211 monitor  $\rightarrow$  O<sub>3</sub>





NPL D180545 NO cylinder NO: 9.99 +- 0.10 μmol/mol CO<sub>2</sub>: 50.43 +- 0.25 mmol/mol

# Audit plan

2024	2025		2026+
Jungfraujoch → Nov.	Pallas	→ Spring	6 – 8 NFs
	SMEAR II → Spring		
	SIRTA	→ Summer	
	NN	→ Autumn	



